



SEDEX Low Temperature Evaporative Light Scattering Detector

SAMPLE: POLYOXYETHYLENE ALCOHOLS
(C₁₆E₇)

HPLC
Gradient



CONDITIONS

Column	<i>Lichrospher 100 Diol - 5 μm</i> <i>125 x 4 mm</i>
Injected Volume	<i>20 μl (Injected Quantity 10 μg)</i>
Eluent	<i>A: Hexane</i> <i>B: Chloroform/Isopropyl alcohol (2%)</i>
Flow Rate	<i>1 ml/min</i>
Gas	<i>Nitrogen</i>
Pressure	<i>2.5 Bars</i>
Temperature	<i>35° C</i>

Analyzed By : Mr. Becart, Parfums C.Dior