


SEDEX Low Temperature Evaporative Light Scattering Detector

SAMPLE: POLYETHYLENE GLYCOLS PEG 600

HPLC
Gradient



Gradient: Acetonitrile/Water 10/90 to 90/10 in 40 min

CONDITIONS

Column	PLRPS - 5 μ m 150 x 4.6 mm
Injected Volume	20 μ l (Injected Quantity 10 μ g)
Eluent	A: Acetonitrile B: Water
Flow Rate	1.5 ml/min
Gas	Air
Pressure	2.5 Bars
Temperature	40° C

Analyzed By : LCBA 45100 Orleans FRANCE