


# SEDEX Low Temperature Evaporative Light Scattering Detector

SAMPLE: ARACHIDONIC D8 ACID

HPLC



## CONDITIONS

Column	<i>Zorbax ODS - 5 <math>\mu</math>m</i> <i>15 x 4.6 mm</i>
Injected Volume	<i>20 <math>\mu</math>l</i>
Eluent	<i>Acetone/MeOH/Water/5% Acetic Acid</i> <i>(40/5/5/50)</i>
Flow Rate	<i>1.5 ml/min</i>
Pressure	<i>2.5 Bars</i>
Gas	<i>Nitrogen</i>
Temperature	<i>40° C</i>

**S.E.D.E.R.E.**

PARC VOLTA - BP 27

9, rue Parmentier

94141 Alfortville Cedex FRANCE

Application #31